

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731
 Lab File ID: VX045586.D Date Analyzed: 04/04/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:00
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	67379	12.018				
UPPER LIMIT	134758	12.518				
LOWER LIMIT	33689.5	11.518				
EPA SAMPLE NO.						
EB01-040325	49367	12.02				
TB01-040325	43047	12.02				
VX0404WBL01	46025	12.02				
VX0404WBS01	66584	12.02				
VX0404WBSD01	65264	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.