

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q1731 SAS No.: Q1731 SDG NO.: Q1731
 Lab File ID: VX045614.D Date Analyzed: 04/07/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:17
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	73594	12.018				
UPPER LIMIT	147188	12.518				
LOWER LIMIT	36797	11.518				
EPA SAMPLE NO.						
RMW-01B-82-040325	42386	12.02				
RMW-04B-91-040325	42372	12.02				
RMW-01B-82-040325-FD	47624	12.02				
RMW-03B-90-040325	48503	12.02				
VX0407WBL01	49926	12.02				
VX0407WBS01	64861	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.