

Continuing Calibration Report for SequenceID : FC040725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068497.D

Aliphatic C9-C12	6682131.000	60.000	3.349	6.652	111368.850	113656.014	2.012
Aliphatic C12-C16	4492250.000	40.000	6.653	10.053	112306.250	112966.617	0.585
Aliphatic C16-C21	6578642.000	60.000	10.054	13.420	109644.033	110233.878	0.535
Aliphatic C21-C28	8495195.000	80.000	13.421	17.087	106189.938	110346.279	3.767
Aliphatic C28-C40	9408552.000	120.000	17.088	22.097	78404.600	81423.284	3.707
Aliphatic EPH	35656770.000	360.000	3.349	22.097	99046.583	101529.318	2.445