

Continuing Calibration Report for SequenceID : FC040725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068506.D

Aliphatic C9-C12	7148989.000	60.000	3.349	6.652	119149.817	113656.014	-4.834
Aliphatic C12-C16	4796745.000	40.000	6.653	10.053	119918.625	112966.617	-6.154
Aliphatic C16-C21	6987395.000	60.000	10.054	13.420	116456.583	110233.878	-5.645
Aliphatic C21-C28	8925778.000	80.000	13.421	17.087	111572.225	110346.279	-1.111
Aliphatic C28-C40	9592709.000	120.000	17.088	22.097	79939.242	81423.284	1.823
Aliphatic EPH	37451616.000	360.000	3.349	22.097	104032.267	101529.318	-2.465