

Continuing Calibration Report for SequenceID : FC040825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068509.D

Aliphatic C9-C12	7037737.000	60.000	3.349	6.653	117295.617	113656.014	-3.202
Aliphatic C12-C16	4773823.000	40.000	6.654	10.055	119345.575	112966.617	-5.647
Aliphatic C16-C21	7017670.000	60.000	10.056	13.422	116961.167	110233.878	-6.103
Aliphatic C21-C28	9090037.000	80.000	13.423	17.087	113625.463	110346.279	-2.972
Aliphatic C28-C40	9806940.000	120.000	17.088	22.094	81724.500	81423.284	-0.370
Aliphatic EPH	37726207.000	360.000	3.349	22.094	104795.019	101529.318	-3.217