

Continuing Calibration Report for SequenceID : FC040825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068512.D

Aliphatic C9-C12	7222772.000	60.000	3.349	6.653	120379.533	113656.014	-5.916
Aliphatic C12-C16	4679529.000	40.000	6.654	10.055	116988.225	112966.617	-3.560
Aliphatic C16-C21	6686553.000	60.000	10.056	13.422	111442.550	110233.878	-1.096
Aliphatic C21-C28	8623503.000	80.000	13.423	17.087	107793.788	110346.279	2.313
Aliphatic C28-C40	9762172.000	120.000	17.088	22.094	81351.433	81423.284	0.088
Aliphatic EPH	36974529.000	360.000	3.349	22.094	102707.025	101529.318	-1.160