

### CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q1736 SAS No.: Q1736 SDG NO.: Q1736

Continuing Calib Date: 04/07/2025 Initial Calibration Date(s): 03/18/2025 03/18/2025

Continuing Calib Time: 22:22 Initial Calibration Time(s): 14:03 22:15

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1242-1       | (1) | 4.78       | 4.79      | 4.69      | 4.89 | 0.01       |
| Aroclor-1242-2       | (2) | 4.80       | 4.81      | 4.71      | 4.91 | 0.01       |
| Aroclor-1242-3       | (3) | 4.86       | 4.86      | 4.76      | 4.96 | 0.00       |
| Aroclor-1242-4       | (4) | 4.98       | 4.98      | 4.88      | 5.08 | 0.00       |
| Aroclor-1242-5       | (5) | 5.63       | 5.64      | 5.54      | 5.74 | 0.01       |
| Tetrachloro-m-xylene |     | 3.69       | 3.69      | 3.59      | 3.79 | 0.00       |
| Decachlorobiphenyl   |     | 8.74       | 8.74      | 8.64      | 8.84 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q1736 SAS No.: Q1736 SDG NO.: Q1736

Continuing Calib Date: 04/07/2025 Initial Calibration Date(s): 03/18/2025 03/18/2025

Continuing Calib Time: 22:22 Initial Calibration Time(s): 14:03 22:15

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1242-1       | (1) | 4.77       | 4.77      | 4.67      | 4.87 | 0.00       |
| Aroclor-1242-2       | (2) | 4.79       | 4.79      | 4.69      | 4.89 | 0.00       |
| Aroclor-1242-3       | (3) | 4.96       | 4.97      | 4.87      | 5.07 | 0.01       |
| Aroclor-1242-4       | (4) | 5.05       | 5.05      | 4.95      | 5.15 | 0.00       |
| Aroclor-1242-5       | (5) | 5.57       | 5.57      | 5.47      | 5.67 | 0.00       |
| Tetrachloro-m-xylene |     | 3.69       | 3.69      | 3.59      | 3.79 | 0.00       |
| Decachlorobiphenyl   |     | 8.69       | 8.70      | 8.60      | 8.80 | 0.01       |