

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q1736 SAS No.: Q1736 SDG NO.: Q1736

Continuing Calib Date: 04/07/2025 Initial Calibration Date(s): 03/18/2025 03/18/2025

Continuing Calib Time: 22:59 Initial Calibration Time(s): 14:03 22:15

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.59	5.59	5.49	5.69	0.00
Aroclor-1254-2 (2)	5.74	5.74	5.64	5.84	0.00
Aroclor-1254-3 (3)	6.15	6.15	6.05	6.25	0.01
Aroclor-1254-4 (4)	6.38	6.38	6.28	6.48	0.00
Aroclor-1254-5 (5)	6.80	6.80	6.70	6.90	0.00
Tetrachloro-m-xylene	3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl	8.74	8.74	8.64	8.84	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q1736 SAS No.: Q1736 SDG NO.: Q1736

Continuing Calib Date: 04/07/2025 Initial Calibration Date(s): 03/18/2025 03/18/2025

Continuing Calib Time: 22:59 Initial Calibration Time(s): 14:03 22:15

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.57	5.57	5.47	5.67	0.00
Aroclor-1254-2 (2)	5.72	5.72	5.62	5.82	0.00
Aroclor-1254-3 (3)	6.12	6.12	6.02	6.22	0.00
Aroclor-1254-4 (4)	6.35	6.35	6.25	6.45	0.00
Aroclor-1254-5 (5)	6.77	6.77	6.67	6.87	0.01
Tetrachloro-m-xylene	3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl	8.69	8.70	8.60	8.80	0.01