

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q1736 SAS No.: Q1736 SDG NO.: Q1736

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL02 Date Analyzed: 04/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO110278.D Time Analyzed: 17:14

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.782	4.685	4.885	511.900	500.000	2.4
Aroclor-1016-2	4.802	4.704	4.904	513.800	500.000	2.8
Aroclor-1016-3	4.858	4.761	4.961	513.560	500.000	2.7
Aroclor-1016-4	4.979	4.881	5.081	505.500	500.000	1.1
Aroclor-1016-5	5.236	5.139	5.339	483.610	500.000	-3.3
Aroclor-1260-1	6.276	6.180	6.380	516.970	500.000	3.4
Aroclor-1260-2	6.466	6.369	6.569	486.490	500.000	-2.7
Aroclor-1260-3	6.834	6.737	6.937	450.780	500.000	-9.8
Aroclor-1260-4	7.094	6.998	7.198	442.100	500.000	-11.6
Aroclor-1260-5	7.336	7.239	7.439	429.820	500.000	-14.0
Decachlorobiphenyl	8.741	8.645	8.845	52.440	50.000	4.9
Tetrachloro-m-xylene	3.691	3.592	3.792	52.070	50.000	4.1

CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q1736 SAS No.: Q1736 SDG NO.: Q1736

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL02 Date Analyzed: 04/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO110278.D Time Analyzed: 17:14

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.769	4.672	4.872	501.230	500.000	0.2
Aroclor-1016-2	4.788	4.691	4.891	505.970	500.000	1.2
Aroclor-1016-3	4.964	4.867	5.067	506.650	500.000	1.3
Aroclor-1016-4	5.006	4.909	5.109	501.000	500.000	0.2
Aroclor-1016-5	5.218	5.122	5.322	500.710	500.000	0.1
Aroclor-1260-1	6.250	6.155	6.355	459.750	500.000	-8.1
Aroclor-1260-2	6.438	6.342	6.542	450.220	500.000	-10.0
Aroclor-1260-3	6.591	6.495	6.695	435.550	500.000	-12.9
Aroclor-1260-4	7.063	6.967	7.167	432.750	500.000	-13.5
Aroclor-1260-5	7.303	7.207	7.407	419.150	500.000	-16.2
Decachlorobiphenyl	8.691	8.596	8.796	46.440	50.000	-7.1
Tetrachloro-m-xylene	3.688	3.590	3.790	52.230	50.000	4.5