

### CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q1736 SAS No.: Q1736 SDG NO.: Q1736

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL03 Date Analyzed: 04/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO110286.D Time Analyzed: 21:27

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Aroclor-1016-1       | 4.783 | 4.685                | 4.885 | 528.320            | 500.000           | 5.7  |
| Aroclor-1016-2       | 4.802 | 4.704                | 4.904 | 526.450            | 500.000           | 5.3  |
| Aroclor-1016-3       | 4.858 | 4.761                | 4.961 | 529.370            | 500.000           | 5.9  |
| Aroclor-1016-4       | 4.979 | 4.881                | 5.081 | 518.990            | 500.000           | 3.8  |
| Aroclor-1016-5       | 5.236 | 5.139                | 5.339 | 496.090            | 500.000           | -0.8 |
| Aroclor-1260-1       | 6.277 | 6.180                | 6.380 | 550.160            | 500.000           | 10.0 |
| Aroclor-1260-2       | 6.466 | 6.369                | 6.569 | 508.400            | 500.000           | 1.7  |
| Aroclor-1260-3       | 6.834 | 6.737                | 6.937 | 481.740            | 500.000           | -3.7 |
| Aroclor-1260-4       | 7.094 | 6.998                | 7.198 | 471.490            | 500.000           | -5.7 |
| Aroclor-1260-5       | 7.336 | 7.239                | 7.439 | 462.920            | 500.000           | -7.4 |
| Decachlorobiphenyl   | 8.740 | 8.645                | 8.845 | 54.690             | 50.000            | 9.4  |
| Tetrachloro-m-xylene | 3.691 | 3.592                | 3.792 | 53.480             | 50.000            | 7.0  |

### CALIBRATION VERIFICATION SUMMARY

Contract: GENV01

Lab Code: CHEM Case No.: Q1736 SAS No.: Q1736 SDG NO.: Q1736

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL03 Date Analyzed: 04/07/2025

Lab Sample No.: AR1660CCC500 Data File : PO110286.D Time Analyzed: 21:27

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|----------------------|-------|--------------------|-------------------|-------|
| Aroclor-1016-1       | 4.769 | 4.672                | 4.872 | 512.420            | 500.000           | 2.5   |
| Aroclor-1016-2       | 4.788 | 4.691                | 4.891 | 514.070            | 500.000           | 2.8   |
| Aroclor-1016-3       | 4.964 | 4.867                | 5.067 | 515.160            | 500.000           | 3.0   |
| Aroclor-1016-4       | 5.006 | 4.909                | 5.109 | 469.250            | 500.000           | -6.2  |
| Aroclor-1016-5       | 5.218 | 5.122                | 5.322 | 514.370            | 500.000           | 2.9   |
| Aroclor-1260-1       | 6.251 | 6.155                | 6.355 | 483.370            | 500.000           | -3.3  |
| Aroclor-1260-2       | 6.439 | 6.342                | 6.542 | 477.180            | 500.000           | -4.6  |
| Aroclor-1260-3       | 6.591 | 6.495                | 6.695 | 462.160            | 500.000           | -7.6  |
| Aroclor-1260-4       | 7.063 | 6.967                | 7.167 | 461.970            | 500.000           | -7.6  |
| Aroclor-1260-5       | 7.303 | 7.207                | 7.407 | 444.510            | 500.000           | -11.1 |
| Decachlorobiphenyl   | 8.691 | 8.596                | 8.796 | 48.770             | 50.000            | -2.5  |
| Tetrachloro-m-xylene | 3.688 | 3.590                | 3.790 | 53.150             | 50.000            | 6.3   |