

Surrogate Summary

SW-846

SDG No.: Q1736

Client: G Environmental

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167475BL	PB167475BL	2-Methylnaphthalene-d10	0.4	0.33	82		30 (17)	150 (161)
		Fluoranthene-d10	0.4	0.41	103		30 (23)	150 (138)
		Nitrobenzene-d5	0.4	0.31	76		30 (33)	130 (121)
		2-Fluorobiphenyl	0.4	0.26	65		30 (32)	130 (121)
		Terphenyl-d14	0.4	0.35	88		30 (21)	130 (130)
PB167475BS	PB167475BS	2-Methylnaphthalene-d10	0.4	0.37	93		30 (17)	150 (161)
		Fluoranthene-d10	0.4	0.36	89		30 (23)	150 (138)
		Nitrobenzene-d5	0.4	0.31	78		30 (33)	130 (121)
		2-Fluorobiphenyl	0.4	0.34	85		30 (32)	130 (121)
		Terphenyl-d14	0.4	0.33	83		30 (21)	130 (130)
Q1736-01	GST1	2-Methylnaphthalene-d10	0.4	0.24	60		30 (17)	150 (161)
		Fluoranthene-d10	0.4	0.31	77		30 (23)	150 (138)
		Nitrobenzene-d5	0.4	0.26	65		30 (33)	130 (121)
		2-Fluorobiphenyl	0.4	0.20	50		30 (32)	130 (121)
		Terphenyl-d14	0.4	0.25	63		30 (21)	130 (130)
Q1736-02	GST2	2-Methylnaphthalene-d10	0.4	0.30	75		30 (17)	150 (161)
		Fluoranthene-d10	0.4	0.38	94		30 (23)	150 (138)
		Nitrobenzene-d5	0.4	0.29	73		30 (33)	130 (121)
		2-Fluorobiphenyl	0.4	0.30	76		30 (32)	130 (121)
		Terphenyl-d14	0.4	0.29	72		30 (21)	130 (130)
Q1736-02MS	GST2MS	2-Methylnaphthalene-d10	0.4	0.27	67		30 (17)	150 (161)
		Fluoranthene-d10	0.4	0.31	78		30 (23)	150 (138)
		Nitrobenzene-d5	0.4	0.25	63		30 (33)	130 (121)
		2-Fluorobiphenyl	0.4	0.27	66		30 (32)	130 (121)
		Terphenyl-d14	0.4	0.26	64		30 (21)	130 (130)
Q1736-02MSD	GST2MSD	2-Methylnaphthalene-d10	0.4	0.30	75		30 (17)	150 (161)
		Fluoranthene-d10	0.4	0.34	84		30 (23)	150 (138)
		Nitrobenzene-d5	0.4	0.28	69		30 (33)	130 (121)
		2-Fluorobiphenyl	0.4	0.27	66		30 (32)	130 (121)
		Terphenyl-d14	0.4	0.28	70		30 (21)	130 (130)