

**ALLIANCE** 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

**METALS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: Q1739

MATRIX: Water

METHOD: 6010D,7470A

	NA	NO	YES
1. Calibration Summary met criteria.			✓
2. ICP Interference Check Sample Results Summary Submitted.			✓
3. Serial Dilution Summary (if applicable) Submitted.			✓
4. Laboratory Control Sample Summary (if applicable) Submitted.			✓
5. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
6. Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria		✓	

If not met, list those compounds and their recoveries which fall outside the acceptable range.

The Matrix Spike (B-158-GW01MS) analysis met criteria for all samples except for Mercury due to matrix interference.

The Matrix Spike (WATER TREATMENT DISCHARGEMS) analysis met criteria for all samples except for Antimony, Barium, Beryllium, Magnesium due to matrix interference. The Matrix Spike Duplicate (B-158-GW01MSD) analysis met criteria for all samples except for Mercury due to matrix interference.

The Matrix Spike Duplicate (WATER TREATMENT DISCHARGEMSD) analysis met criteria for all samples except for Antimony, Barium, Beryllium, Magnesium due to matrix interference.

7. Sample Duplicate Analysis Met QC Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
8. Digestion Holding Time Met			✓
If not met, list number of days exceeded for each sample:			
9. Analysis Holding Time Met			✓

If not met, list those compounds and their recoveries which fall outside the acceptable range.

ADDITIONAL COMMENTS:

\_\_\_\_\_  
QA REVIEW

\_\_\_\_\_  
Date