



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q1739

Project: Con Edison - 11th Ave-West 50th St Site

RunNo.: LB135330

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.00	7	100	90-110	04/07/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	04/07/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	04/07/2025

Initial and Continuing Calibration Verification

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q1739

Project: Con Edison - 11th Ave-West 50th St Site

RunNo.: LB135340

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV						
Flash Point	o F	82.4	81	102	78-84	04/08/2025

Initial and Continuing Calibration Verification

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q1739

Project: Con Edison - 11th Ave-West 50th St Site

RunNo.: LB135341

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.09	0.099	91	85-115	04/08/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	04/08/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	04/08/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	04/08/2025