



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q1739

Contract: PARS02

Lab Code: CHEM

Case No.: Q1739

SAS No.: Q1739

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB84	Mercury	0.20	+/-0.20	U	0.20	CV	04/11/2025	10:05	LB135388

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q1739

Contract: PARS02

Lab Code: CHEM

Case No.: Q1739

SAS No.: Q1739

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB59	Mercury	0.20	+/-0.20	U	0.20	CV	04/11/2025	10:09	LB135388
CCB60	Mercury	0.20	+/-0.20	U	0.20	CV	04/11/2025	10:39	LB135388
CCB61	Mercury	0.20	+/-0.20	U	0.20	CV	04/11/2025	11:16	LB135388
CCB62	Mercury	0.20	+/-0.20	U	0.20	CV	04/11/2025	11:43	LB135388

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q1739

Contract: PARS02

Lab Code: CHEM

Case No.: Q1739

SAS No.: Q1739

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	04/10/2025	10:29	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	10:29	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	10:29	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	10:29	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	10:29	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	10:29	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	10:29	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	10:29	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	10:29	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	10:29	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	10:29	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	10:29	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	10:29	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	10:29	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	10:29	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	10:29	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	10:29	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	10:29	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	10:29	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	10:29	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	10:29	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	10:29	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q1739

Contract: PARS02

Lab Code: CHEM

Case No.: Q1739

SAS No.: Q1739

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	04/10/2025	11:21	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	11:21	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	11:21	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	11:21	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	11:21	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	11:21	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	11:21	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	11:21	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	11:21	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	11:21	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	11:21	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	11:21	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	11:21	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	11:21	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	11:21	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	11:21	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	11:21	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	11:21	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	11:21	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	11:21	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	11:21	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	11:21	LB135373
CCB02	Aluminum	100	+/-100	U	100	P	04/10/2025	12:12	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	12:12	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	12:12	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	12:12	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	12:12	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	12:12	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	12:12	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	12:12	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	12:12	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	12:12	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	12:12	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	12:12	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	12:12	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	12:12	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	12:12	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	12:12	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	12:12	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q1739

Contract: PARS02

Lab Code: CHEM

Case No.: Q1739

SAS No.: Q1739

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	12:12	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	12:12	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	12:12	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	12:12	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	12:12	LB135373
CCB03	Aluminum	100	+/-100	U	100	P	04/10/2025	13:04	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	13:04	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	13:04	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	13:04	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	13:04	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	13:04	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	13:04	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	13:04	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	13:04	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	13:04	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	13:04	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	13:04	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	13:04	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	13:04	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	13:04	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	13:04	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	13:04	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	13:04	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	13:04	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	13:04	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	13:04	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	13:04	LB135373
CCB04	Aluminum	100	+/-100	U	100	P	04/10/2025	13:55	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	13:55	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	13:55	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	13:55	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	13:55	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	13:55	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	13:55	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	13:55	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	13:55	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	13:55	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	13:55	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	13:55	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q1739

Contract: PARS02

Lab Code: CHEM

Case No.: Q1739

SAS No.: Q1739

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	13:55	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	13:55	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	13:55	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	13:55	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	13:55	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	13:55	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	13:55	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	13:55	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	13:55	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	13:55	LB135373
CCB05	Aluminum	100	+/-100	U	100	P	04/10/2025	14:49	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	14:49	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	14:49	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	14:49	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	14:49	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	14:49	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	14:49	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	14:49	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	14:49	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	14:49	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	14:49	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	14:49	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	14:49	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	14:49	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	14:49	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	14:49	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	14:49	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	14:49	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	14:49	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	14:49	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	14:49	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	14:49	LB135373
CCB06	Aluminum	100	+/-100	U	100	P	04/10/2025	15:48	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	15:48	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	15:48	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	15:48	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	15:48	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	15:48	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	15:48	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q1739

Contract: PARS02

Lab Code: CHEM

Case No.: Q1739

SAS No.: Q1739

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	15:48	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	15:48	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	15:48	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	15:48	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	15:48	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	15:48	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	15:48	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	15:48	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	15:48	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	15:48	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	15:48	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	15:48	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	15:48	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	15:48	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	15:48	LB135373
CCB07	Aluminum	100	+/-100	U	100	P	04/10/2025	17:00	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	17:00	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	17:00	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	17:00	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	17:00	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	17:00	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	17:00	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	17:00	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	17:00	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	17:00	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	17:00	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	17:00	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	17:00	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	17:00	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	17:00	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	17:00	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	17:00	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	17:00	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	17:00	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	17:00	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	17:00	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	17:00	LB135373
CCB08	Aluminum	100	+/-100	U	100	P	04/10/2025	18:00	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	18:00	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q1739

Contract: PARS02

Lab Code: CHEM

Case No.: Q1739

SAS No.: Q1739

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	18:00	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	18:00	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	18:00	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	18:00	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	18:00	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	18:00	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	18:00	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	18:00	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	18:00	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	18:00	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	18:00	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	18:00	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	18:00	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	18:00	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	18:00	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	18:00	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	18:00	LB135373
	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	18:00	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	18:00	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	18:00	LB135373
CCB09	Aluminum	100	+/-100	U	100	P	04/10/2025	18:23	LB135373
	Antimony	50.0	+/-50.0	U	50.0	P	04/10/2025	18:23	LB135373
	Arsenic	20.0	+/-20.0	U	20.0	P	04/10/2025	18:23	LB135373
	Barium	100	+/-100	U	100	P	04/10/2025	18:23	LB135373
	Beryllium	6.00	+/-6.00	U	6.00	P	04/10/2025	18:23	LB135373
	Cadmium	6.00	+/-6.00	U	6.00	P	04/10/2025	18:23	LB135373
	Calcium	2000	+/-2000	U	2000	P	04/10/2025	18:23	LB135373
	Chromium	10.0	+/-10.0	U	10.0	P	04/10/2025	18:23	LB135373
	Cobalt	30.0	+/-30.0	U	30.0	P	04/10/2025	18:23	LB135373
	Copper	20.0	+/-20.0	U	20.0	P	04/10/2025	18:23	LB135373
	Iron	100	+/-100	U	100	P	04/10/2025	18:23	LB135373
	Lead	12.0	+/-12.0	U	12.0	P	04/10/2025	18:23	LB135373
	Magnesium	2000	+/-2000	U	2000	P	04/10/2025	18:23	LB135373
	Manganese	20.0	+/-20.0	U	20.0	P	04/10/2025	18:23	LB135373
	Nickel	40.0	+/-40.0	U	40.0	P	04/10/2025	18:23	LB135373
	Potassium	2000	+/-2000	U	2000	P	04/10/2025	18:23	LB135373
	Selenium	20.0	+/-20.0	U	20.0	P	04/10/2025	18:23	LB135373
	Silver	10.0	+/-10.0	U	10.0	P	04/10/2025	18:23	LB135373
	Sodium	2000	+/-2000	U	2000	P	04/10/2025	18:23	LB135373

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q1739

Contract: PARS02

Lab Code: CHEM

Case No.: Q1739

SAS No.: Q1739

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	04/10/2025	18:23	LB135373
	Vanadium	40.0	+/-40.0	U	40.0	P	04/10/2025	18:23	LB135373
	Zinc	40.0	+/-40.0	U	40.0	P	04/10/2025	18:23	LB135373