



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q1739

Contract: PARS02

Lab Code: CHEM

Case No.: Q1739

SAS No.: Q1739

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	19.0	20.0	95	40 - 160	P	04/08/2025	15:00	LB135348
	Barium	83.4	100	83	40 - 160	P	04/08/2025	15:00	LB135348
	Cadmium	5.70	6.0	95	40 - 160	P	04/08/2025	15:00	LB135348
	Chromium	10.8	10.0	108	40 - 160	P	04/08/2025	15:00	LB135348
	Lead	12.2	12.0	102	40 - 160	P	04/08/2025	15:00	LB135348
	Selenium	21.8	20.0	109	40 - 160	P	04/08/2025	15:00	LB135348
	Silver	11.0	10.0	110	40 - 160	P	04/08/2025	15:00	LB135348
CRA	Mercury	0.21	0.2	104	40 - 160	CV	04/11/2025	10:12	LB135388