

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: PARSONS Engineering of New York, Inc. **SDG No.:** Q1739
Contract: PARS02 **Lab Code:** CHEM **Case No.:** Q1739 **SAS No.:** Q1739
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	2.21			-20	20	04/08/2025	15:04	LB135348
	Barium	-5.93	6.0	99	-94	106	04/08/2025	15:04	LB135348
	Cadmium	1.48	1.0	148	-5	7	04/08/2025	15:04	LB135348
	Chromium	59.5	52.0	114	42	62	04/08/2025	15:04	LB135348
	Lead	4.71			-12	12	04/08/2025	15:04	LB135348
	Selenium	4.79			-20	20	04/08/2025	15:04	LB135348
	Silver	-1.86			-10	10	04/08/2025	15:04	LB135348
ICSAB01	Arsenic	114	104	110	88.4	120	04/08/2025	15:09	LB135348
	Barium	462	537	86	437	637	04/08/2025	15:09	LB135348
	Cadmium	988	972	102	826	1120	04/08/2025	15:09	LB135348
	Chromium	585	542	108	460	624	04/08/2025	15:09	LB135348
	Lead	54.4	49.0	111	37	61	04/08/2025	15:09	LB135348
	Selenium	55.2	46.0	120	26	66	04/08/2025	15:09	LB135348
	Silver	187	201	93	170	232	04/08/2025	15:09	LB135348