

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: PARSONS Engineering of New York, Inc. **SDG No.:** Q1739
Contract: PARS02 **Lab Code:** CHEM **Case No.:** Q1739 **SAS No.:** Q1739
Matrix: Water **Level:** LOW **Client ID:** TP-20A
Sample ID: Q1740-04 **Spiked ID:** Q1740-04A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Barium	ug/L	75 - 125	2960		2120		1000	84		P