

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: PARSONS Engineering of New York, Inc. **Level:** LOW **SDG No.:** Q1739
Contract: PARS02 **Lab Code:** CHEM **Case No.:** Q1739 **SAS No.:** Q1739
Matrix: Water **Sample ID:** Q1739-02 **Client ID:** WC-LIQUID-20250404DUP
Percent Solids for Sample: NA **Duplicate ID** Q1739-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	2.00	U	2.00	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: PARSONS Engineering of New York, Inc. **Level:** LOW **SDG No.:** Q1739
Contract: PARS02 **Lab Code:** CHEM **Case No.:** Q1739 **SAS No.:** Q1739
Matrix: Water **Sample ID:** Q1739-02MS **Client ID:** WC-LIQUID-20250404MSD
Percent Solids for Sample: NA **Duplicate ID** Q1739-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	35.3		38.2		8		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: PARSONS Engineering of New York, Inc. **Level:** LOW **SDG No.:** Q1739
Contract: PARS02 **Lab Code:** CHEM **Case No.:** Q1739 **SAS No.:** Q1739
Matrix: Water **Sample ID:** Q1740-04 **Client ID:** TP-20DUP
Percent Solids for Sample: NA **Duplicate ID** Q1740-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	2120		2650		22	*	P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	55.2	J	66.9		19		P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: PARSONS Engineering of New York, Inc. **Level:** LOW **SDG No.:** Q1739
Contract: PARS02 **Lab Code:** CHEM **Case No.:** Q1739 **SAS No.:** Q1739
Matrix: Water **Sample ID:** Q1740-04MS **Client ID:** TP-20MSD
Percent Solids for Sample: NA **Duplicate ID** Q1740-04MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4180		4430		6		P
Barium	ug/L	20	3010		3580		17		P
Cadmium	ug/L	20	928		943		2		P
Chromium	ug/L	20	2150		2170		1		P
Lead	ug/L	20	4450		4530		2		P
Selenium	ug/L	20	9970		10600		6		P
Silver	ug/L	20	380		392		3		P