

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG NO.: Q1739

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/18/2025 03/18/2025

Continuing Calib Time: 09:45 Initial Calibration Time(s): 14:03 22:15

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.78	4.79	4.69	4.89	0.01
Aroclor-1248-2	(2)	5.02	5.02	4.92	5.12	0.00
Aroclor-1248-3	(3)	5.24	5.24	5.14	5.34	0.00
Aroclor-1248-4	(4)	5.59	5.59	5.49	5.69	0.00
Aroclor-1248-5	(5)	5.63	5.63	5.53	5.73	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.74	8.74	8.64	8.84	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG NO.: Q1739

Continuing Calib Date: 04/08/2025 Initial Calibration Date(s): 03/18/2025 03/18/2025

Continuing Calib Time: 09:45 Initial Calibration Time(s): 14:03 22:15

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1248-2	(2)	5.01	5.01	4.91	5.11	0.00
Aroclor-1248-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1248-4	(4)	5.22	5.22	5.12	5.32	0.00
Aroclor-1248-5	(5)	5.61	5.61	5.51	5.71	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.69	8.70	8.60	8.80	0.01