

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG NO.: Q1739

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL02 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PO110315.D Time Analyzed: 17:54

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.781	4.685	4.885	476.910	500.000	-4.6
Aroclor-1016-2	4.801	4.704	4.904	479.380	500.000	-4.1
Aroclor-1016-3	4.857	4.761	4.961	481.950	500.000	-3.6
Aroclor-1016-4	4.977	4.881	5.081	474.440	500.000	-5.1
Aroclor-1016-5	5.235	5.139	5.339	487.450	500.000	-2.5
Aroclor-1260-1	6.276	6.180	6.380	480.600	500.000	-3.9
Aroclor-1260-2	6.465	6.369	6.569	458.870	500.000	-8.2
Aroclor-1260-3	6.832	6.737	6.937	467.790	500.000	-6.4
Aroclor-1260-4	7.092	6.998	7.198	459.450	500.000	-8.1
Aroclor-1260-5	7.335	7.239	7.439	450.870	500.000	-9.8
Decachlorobiphenyl	8.738	8.645	8.845	55.330	50.000	10.7
Tetrachloro-m-xylene	3.689	3.592	3.792	47.630	50.000	-4.7

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG NO.: Q1739

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL02 Date Analyzed: 04/08/2025

Lab Sample No.: AR1660CCC500 Data File : PO110315.D Time Analyzed: 17:54

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.768	4.672	4.872	459.430	500.000	-8.1
Aroclor-1016-2	4.787	4.691	4.891	459.990	500.000	-8.0
Aroclor-1016-3	4.963	4.867	5.067	461.620	500.000	-7.7
Aroclor-1016-4	5.005	4.909	5.109	457.450	500.000	-8.5
Aroclor-1016-5	5.217	5.122	5.322	464.660	500.000	-7.1
Aroclor-1260-1	6.250	6.155	6.355	444.580	500.000	-11.1
Aroclor-1260-2	6.438	6.342	6.542	437.380	500.000	-12.5
Aroclor-1260-3	6.591	6.495	6.695	422.190	500.000	-15.6
Aroclor-1260-4	7.062	6.967	7.167	419.140	500.000	-16.2
Aroclor-1260-5	7.303	7.207	7.407	409.380	500.000	-18.1
Decachlorobiphenyl	8.691	8.596	8.796	44.500	50.000	-11.0
Tetrachloro-m-xylene	3.687	3.590	3.790	47.340	50.000	-5.3