

Surrogate Summary

SDG No.: Q1739

Client: PARSONS Engineering of New York, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO109971.D	PIBLK-PO109971.D	Tetrachloro-m-xylene	1	20	18.5	93		60	140
		Decachlorobiphenyl	1	20	20.0	100		60	140
		Tetrachloro-m-xylene	2	20	18.6	93		60	140
		Decachlorobiphenyl	2	20	21.3	107		60	140
I.BLK-PO110306.D	PIBLK-PO110306.D	Tetrachloro-m-xylene	1	20	18.2	91		60	140
		Decachlorobiphenyl	1	20	18.4	92		60	140
		Tetrachloro-m-xylene	2	20	17.4	87		60	140
		Decachlorobiphenyl	2	20	15.3	77		60	140
PB167515BL	PB167515BL	Tetrachloro-m-xylene	1	20	19.6	98		16	158
		Decachlorobiphenyl	1	20	19.8	99		10	173
		Tetrachloro-m-xylene	2	20	18.6	93		16	158
		Decachlorobiphenyl	2	20	16.4	82		10	173
PB167515BS	PB167515BS	Tetrachloro-m-xylene	1	20	19.6	98		16	158
		Decachlorobiphenyl	1	20	21.2	106		10	173
		Tetrachloro-m-xylene	2	20	18.6	93		16	158
		Decachlorobiphenyl	2	20	17.3	87		10	173
PB167515BSD	PB167515BSD	Tetrachloro-m-xylene	1	20	19.1	95		16	158
		Decachlorobiphenyl	1	20	20.6	103		10	173
		Tetrachloro-m-xylene	2	20	18.1	91		16	158
		Decachlorobiphenyl	2	20	16.7	83		10	173
Q1739-01	WC-LIQUID-20250404	Tetrachloro-m-xylene	1	20	12.2	61		16	158
		Decachlorobiphenyl	1	20	11.6	58		10	173
		Tetrachloro-m-xylene	2	20	13.0	65		16	158
		Decachlorobiphenyl	2	20	9.60	48		10	173
I.BLK-PO110319.D	PIBLK-PO110319.D	Tetrachloro-m-xylene	1	20	17.8	89		60	140
		Decachlorobiphenyl	1	20	18.2	91		60	140
		Tetrachloro-m-xylene	2	20	17.3	86		60	140
		Decachlorobiphenyl	2	20	15.5	77		60	140