

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM **Case No.:** Q1739 **SAS No.:** Q1739 **SDG NO.:** Q1739

Continuing Calib Date: 04/09/2025 **Initial Calibration Date(s):** 04/02/2025 04/02/2025

Continuing Calib Time: 20:34 **Initial Calibration Time(s):** 17:32 20:44

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	6.94	6.96	6.86	7.06	0.02
2,4-D	8.02	8.04	7.94	8.14	0.02
2,4,5-TP(Silvex)	8.86	8.89	8.79	8.99	0.03

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM **Case No.:** Q1739 **SAS No.:** Q1739 **SDG NO.:** Q1739

Continuing Calib Date: 04/09/2025 **Initial Calibration Date(s):** 04/02/2025 04/02/2025

Continuing Calib Time: 20:34 **Initial Calibration Time(s):** 17:32 20:44

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.47	7.48	7.38	7.58	0.01
2,4-D	8.66	8.67	8.57	8.77	0.01
2,4,5-TP(Silvex)	9.54	9.55	9.45	9.65	0.01