

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1739

Client: PARSONS Engineering of New York, In

Analytical Method: 8151A

DataFile : PS029767.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID:	WC-LIQUID-20250404MSD											
Q1739-02MSD	2,4-D	50	0	60.0	ug/L	120		2		65	135	20
	2,4,5-TP(Silvex)	50	0	62.0	ug/L	124		2		62	139	20