

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG NO.: Q1739

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 04/09/2025

Lab Sample No.: PSTDCCC050 Data File : PL095163.D Time Analyzed: 23:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.045	8.956	9.156	50.730	50.000	1.5
Endrin	6.567	6.474	6.674	41.680	50.000	-16.6
gamma-BHC (Lindane)	4.322	4.227	4.427	46.400	50.000	-7.2
Heptachlor	4.909	4.815	5.015	43.480	50.000	-13.0
Heptachlor epoxide	5.678	5.583	5.783	46.450	50.000	-7.1
Methoxychlor	7.494	7.400	7.600	47.060	50.000	-5.9
Tetrachloro-m-xylene	3.534	3.438	3.638	46.840	50.000	-6.3

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG NO.: Q1739

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 04/09/2025

Lab Sample No.: PSTDCCC050 Data File : PL095163.D Time Analyzed: 23:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.897	7.807	8.007	50.810	50.000	1.6
Endrin	5.625	5.534	5.734	46.200	50.000	-7.6
gamma-BHC (Lindane)	3.599	3.504	3.704	50.930	50.000	1.9
Heptachlor	3.935	3.842	4.042	47.430	50.000	-5.1
Heptachlor epoxide	4.717	4.625	4.825	48.940	50.000	-2.1
Methoxychlor	6.598	6.507	6.707	47.380	50.000	-5.2
Tetrachloro-m-xylene	2.768	2.672	2.872	51.170	50.000	2.3