

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG NO.: Q1739

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL05 Date Analyzed: 04/15/2025

Lab Sample No.: PSTDCCC050 Data File : PL095236.D Time Analyzed: 15:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.063	8.953	9.153	51.570	50.000	3.1
Endrin	6.579	6.471	6.671	55.680	50.000	11.4
gamma-BHC (Lindane)	4.331	4.223	4.423	51.130	50.000	2.3
Heptachlor	4.919	4.811	5.011	52.890	50.000	5.8
Heptachlor epoxide	5.688	5.580	5.780	51.510	50.000	3.0
Methoxychlor	7.507	7.398	7.598	49.330	50.000	-1.3
Tetrachloro-m-xylene	3.543	3.435	3.635	55.220	50.000	10.4

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG NO.: Q1739

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL05 Date Analyzed: 04/15/2025

Lab Sample No.: PSTDCCC050 Data File : PL095236.D Time Analyzed: 15:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.906	7.800	8.000	43.400	50.000	-13.2
Endrin	5.631	5.527	5.727	53.150	50.000	6.3
gamma-BHC (Lindane)	3.602	3.499	3.699	49.310	50.000	-1.4
Heptachlor	3.940	3.836	4.036	52.310	50.000	4.6
Heptachlor epoxide	4.722	4.618	4.818	51.290	50.000	2.6
Methoxychlor	6.605	6.499	6.699	50.080	50.000	0.2
Tetrachloro-m-xylene	2.770	2.667	2.867	52.570	50.000	5.1