

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG NO.: Q1739

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No. (PEM): PEM - PL095128.D Date Analyzed: 04/09/2025

Lab Sample No.(PEM): PEM Time Analyzed: 08:43

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.049	8.950	9.150	26.150	20.000	30.8
Tetrachloro-m-xylene	3.536	3.490	3.590	23.490	20.000	17.5
alpha-BHC	3.992	3.940	4.040	12.010	10.000	20.1
beta-BHC	4.523	4.470	4.570	12.040	10.000	20.4
gamma-BHC (Lindane)	4.324	4.270	4.370	12.060	10.000	20.6
Endrin	6.569	6.500	6.640	50.710	50.000	1.4
4,4'-DDT	7.020	6.950	7.090	112.470	100.000	12.5
Methoxychlor	7.497	7.430	7.570	271.750	250.000	8.7

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No. (PEM): PEM - PL095128.D Date Analyzed: 04/09/2025

Lab Sample No.(PEM): PEM Time Analyzed: 08:43

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	7.900	7.800	8.000	24.640	20.000	23.2
Tetrachloro-m-xylene	2.770	2.720	2.820	24.610	20.000	23.1
alpha-BHC	3.271	3.220	3.320	11.740	10.000	17.4
beta-BHC	3.901	3.850	3.950	13.350	10.000	33.5
gamma-BHC (Lindane)	3.600	3.550	3.650	11.860	10.000	18.6
Endrin	5.627	5.560	5.700	56.090	50.000	12.2
4,4'-DDT	6.026	5.960	6.100	122.140	100.000	22.1
Methoxychlor	6.599	6.530	6.670	276.490	250.000	10.6