

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031125

Review By	Abdul	Review On	3/12/2025 12:48:59 PM
Supervise By	Ankita	Supervise On	3/12/2025 2:35:44 PM
SubDirectory	PL031125	HP Acquire Method	HP Processing Method p031125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23793,PP24095
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL094565.D	11 Mar 2025 09:41		AR\AJ	Ok
2	I.BLK	I.BLK	PL094566.D	11 Mar 2025 09:55		AR\AJ	Ok
3	PEM	PEM	PL094567.D	11 Mar 2025 10:08		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL094568.D	11 Mar 2025 10:22		AR\AJ	Ok,M
5	PSTDICC100	PSTDICC100	PL094569.D	11 Mar 2025 10:35		AR\AJ	Ok,M
6	PSTDICC075	PSTDICC075	PL094570.D	11 Mar 2025 10:49		AR\AJ	Ok,M
7	PSTDICC050	PSTDICC050	PL094571.D	11 Mar 2025 11:02		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL094572.D	11 Mar 2025 11:16		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PL094573.D	11 Mar 2025 11:29		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL094574.D	11 Mar 2025 11:43		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL094575.D	11 Mar 2025 11:57		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL094576.D	11 Mar 2025 12:10		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL094577.D	11 Mar 2025 12:24		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL094578.D	11 Mar 2025 12:37		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PL094579.D	11 Mar 2025 12:51		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL094580.D	11 Mar 2025 13:04		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL094581.D	11 Mar 2025 13:18		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL094582.D	11 Mar 2025 13:31		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031125

Review By	Abdul	Review On	3/12/2025 12:48:59 PM
Supervise By	Ankita	Supervise On	3/12/2025 2:35:44 PM
SubDirectory	PL031125	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL094583.D	11 Mar 2025 13:45		AR\AJ	Ok,M
20	PSTDICV050	ICVPL031125	PL094584.D	11 Mar 2025 13:59		AR\AJ	Ok
21	PCHLORICV500	ICVPL031125CHLOR	PL094585.D	11 Mar 2025 14:26		AR\AJ	Ok
22	PTOXICV500	ICVPL031125TOX	PL094586.D	11 Mar 2025 14:53		AR\AJ	Ok
23	I.BLK	I.BLK	PL094587.D	11 Mar 2025 17:16		AR\AJ	Ok
24	PEM	PEM	PL094588.D	11 Mar 2025 17:30		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PL094589.D	11 Mar 2025 17:43		AR\AJ	Ok,M
26	PB167076BL	PB167076BL	PL094590.D	11 Mar 2025 17:57		AR\AJ	Ok
27	PB167076BS	PB167076BS	PL094591.D	11 Mar 2025 18:11		AR\AJ	Ok,M
28	PB167076BSD	PB167076BSD	PL094592.D	11 Mar 2025 18:44		AR\AJ	Ok,M
29	Q1494-01	PURGE-WATER	PL094593.D	11 Mar 2025 18:57	F Flag in TCMX for both column	AR\AJ	Not Ok
30	Q1502-11	PT-CHLR-WP	PL094594.D	11 Mar 2025 19:11	Chlordane CCAL missing , TCMX high in 2nd column	AR\AJ	Not Ok
31	Q1502-09	PT-PEST-WP	PL094595.D	11 Mar 2025 19:25	TCMX high in 1st column , Need dilution	AR\AJ	Dilution
32	Q1502-13	PT-TXP-WP	PL094596.D	11 Mar 2025 19:39	TOX CCAL missing , need dilution	AR\AJ	Not Ok
33	I.BLK	I.BLK	PL094597.D	11 Mar 2025 19:52		AR\AJ	Ok
34	PSTDCCC050	PSTDCCC050	PL094598.D	11 Mar 2025 20:06		AR\AJ	Ok,M
35	Q1539-01	TAPIAL3-MW03D-0310	PL094599.D	11 Mar 2025 20:33		AR\AJ	Ok,M
36	Q1539-02	TAPFTA-MW01I-03102	PL094600.D	11 Mar 2025 20:47		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031125

Review By	Abdul	Review On	3/12/2025 12:48:59 PM
Supervise By	Ankita	Supervise On	3/12/2025 2:35:44 PM
SubDirectory	PL031125	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	PB167086BL	PB167086BL	PL094601.D	11 Mar 2025 21:01		AR\AJ	Ok
38	PB167086BS	PB167086BS	PL094602.D	11 Mar 2025 21:14	Chlordane CCAL missing	AR\AJ	Not Ok
39	PB167087BL	PB167087BL	PL094603.D	11 Mar 2025 21:28		AR\AJ	Ok
40	PB167087BS	PB167087BS	PL094604.D	11 Mar 2025 21:42	TOX CCAL missing	AR\AJ	Not Ok
41	I.BLK	I.BLK	PL094605.D	11 Mar 2025 21:55		AR\AJ	Ok
42	PEM	PEM	PL094606.D	11 Mar 2025 22:09		AR\AJ	Ok,M
43	PSTDCCC050	PSTDCCC050	PL094607.D	11 Mar 2025 22:23		AR\AJ	Ok,M
44	PB167077BL	PB167077BL	PL094608.D	11 Mar 2025 22:50		AR\AJ	Ok
45	PB167077BS	PB167077BS	PL094609.D	11 Mar 2025 23:04	Comp#2 recovery fail	AR\AJ	Not Ok
46	Q1534-01	OR-636-COMP-16	PL094610.D	11 Mar 2025 23:17		AR\AJ	Ok,M
47	Q1534-07	OR-636-COMP-17	PL094611.D	11 Mar 2025 23:31		AR\AJ	Ok,M
48	Q1534-07MS	OR-636-COMP-17MS	PL094612.D	11 Mar 2025 23:44		AR\AJ	Ok,M
49	Q1534-07MSD	OR-636-COMP-17MSD	PL094613.D	11 Mar 2025 23:58		AR\AJ	Ok,M
50	Q1534-13	OR-636-COMP-18	PL094614.D	12 Mar 2025 00:11	need dilution	AR\AJ	Dilution
51	Q1534-19	OR-636-COMP-19	PL094615.D	12 Mar 2025 00:25	need dilution	AR\AJ	Dilution
52	Q1535-01	SU-03-03102025	PL094616.D	12 Mar 2025 00:39		AR\AJ	Ok,M
53	I.BLK	I.BLK	PL094617.D	12 Mar 2025 00:53		AR\AJ	Ok
54	PSTDCCC050	PSTDCCC050	PL094618.D	12 Mar 2025 01:06		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL040925

Review By	Abdul	Review On	4/10/2025 9:04:04 AM
Supervise By	mohammad	Supervise On	4/11/2025 1:23:20 AM
SubDirectory	PL040925	HP Acquire Method	HP Processing Method p031125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24095
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL095126.D	09 Apr 2025 08:16		AR\AJ	Ok
2	I.BLK	I.BLK	PL095127.D	09 Apr 2025 08:29		AR\AJ	Ok
3	PEM	PEM	PL095128.D	09 Apr 2025 08:43		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL095129.D	09 Apr 2025 08:57		AR\AJ	Ok,M
5	PB167473BS	PB167473BS	PL095130.D	09 Apr 2025 11:03		AR\AJ	Ok,M
6	PB167523BL	PB167523BL	PL095131.D	09 Apr 2025 12:28		AR\AJ	Ok
7	PB167523BS	PB167523BS	PL095132.D	09 Apr 2025 14:50		AR\AJ	Ok,M
8	Q1748-01	IB-1.5-WC	PL095133.D	09 Apr 2025 15:41		AR\AJ	Ok,M
9	Q1748-05	IB-2A-WC	PL095134.D	09 Apr 2025 15:54		AR\AJ	Ok,M
10	Q1754-01	TP-1	PL095135.D	09 Apr 2025 16:08		AR\AJ	Ok,M
11	Q1754-03	TP-1-CONCRETE	PL095136.D	09 Apr 2025 16:22		AR\AJ	Ok,M
12	I.BLK	I.BLK	PL095137.D	09 Apr 2025 16:36		AR\AJ	Ok
13	PSTDCCC050	PSTDCCC050	PL095138.D	09 Apr 2025 16:49		AR\AJ	Ok,M
14	Q1749-01	TP-14	PL095139.D	09 Apr 2025 17:36		AR\AJ	Ok,M
15	Q1752-01	TP-1	PL095140.D	09 Apr 2025 17:50		AR\AJ	Ok,M
16	Q1752-03	TP-2	PL095141.D	09 Apr 2025 18:03		AR\AJ	Ok,M
17	Q1752-03MS	TP-2MS	PL095142.D	09 Apr 2025 18:17		AR\AJ	Ok,M
18	Q1752-03MSD	TP-2MSD	PL095143.D	09 Apr 2025 18:31		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL040925

Review By	Abdul	Review On	4/10/2025 9:04:04 AM
Supervise By	mohammad	Supervise On	4/11/2025 1:23:20 AM
SubDirectory	PL040925	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1739-02	WC-LIQUID-20250404	PL095144.D	09 Apr 2025 18:45	TCMX low in both column, confirmed with MS/MSD	AR\AJ	Ok,M
20	Q1739-02MS	WC-LIQUID-20250404	PL095145.D	09 Apr 2025 18:58	TCMX low in both column , Comp#8,14,20 recovery fail	AR\AJ	Ok,M
21	Q1739-02MSD	WC-LIQUID-20250404	PL095146.D	09 Apr 2025 19:12	TCMX low in both column , Comp#8,14,20 recovery fail	AR\AJ	Ok,M
22	I.BLK	I.BLK	PL095147.D	09 Apr 2025 19:26		AR\AJ	Ok
23	PEM	PEM	PL095148.D	09 Apr 2025 19:39		AR\AJ	Ok,M
24	PSTDCCC050	PSTDCCC050	PL095149.D	09 Apr 2025 20:20		AR\AJ	Ok,M
25	Q1752-05	TP-4	PL095150.D	09 Apr 2025 20:34		AR\AJ	Ok,M
26	Q1752-07	TP-3	PL095151.D	09 Apr 2025 20:48		AR\AJ	Ok,M
27	Q1753-01	WC-1	PL095152.D	09 Apr 2025 21:02		AR\AJ	Ok
28	Q1753-05	WC-2	PL095153.D	09 Apr 2025 21:15		AR\AJ	Ok
29	PB167535BL	PB167535BL	PL095154.D	09 Apr 2025 21:29		AR\AJ	Ok
30	PB167535BS	PB167535BS	PL095155.D	09 Apr 2025 21:43	All surrogate fail , Comp#16,20 recovery fail	AR\AJ	Not Ok
31	PB167488TB	PB167488TB	PL095156.D	09 Apr 2025 21:56		AR\AJ	Ok
32	PB167517TB	PB167517TB	PL095157.D	09 Apr 2025 22:10		AR\AJ	Ok
33	Q1744-02	B-158-SB01	PL095158.D	09 Apr 2025 22:24		AR\AJ	Ok,M
34	Q1744-04	B-158-SB02	PL095159.D	09 Apr 2025 22:38		AR\AJ	Ok,M
35	Q1746-02	B-149-SB01	PL095160.D	09 Apr 2025 22:51		AR\AJ	Ok,M
36	Q1746-04	B-149-SB02	PL095161.D	09 Apr 2025 23:05		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL040925

Review By	Abdul	Review On	4/10/2025 9:04:04 AM
Supervise By	mohammad	Supervise On	4/11/2025 1:23:20 AM
SubDirectory	PL040925	HP Acquire Method	HP Processing Method pl031125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24095
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

37	I.BLK	I.BLK	PL095162.D	09 Apr 2025 23:19		AR\AJ	Ok
38	PSTDCCC050	PSTDCCC050	PL095163.D	09 Apr 2025 23:33		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL041425

Review By	Abdul	Review On	4/15/2025 7:51:14 AM
Supervise By	mohammad	Supervise On	4/16/2025 1:08:48 AM
SubDirectory	PL041425	HP Acquire Method	HP Processing Method p041425 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24095
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL095201.D	14 Apr 2025 14:12		AR\AJ	Ok
2	I.BLK	I.BLK	PL095202.D	14 Apr 2025 14:26		AR\AJ	Ok
3	PEM	PEM	PL095203.D	14 Apr 2025 14:40		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL095204.D	14 Apr 2025 14:54		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PL095205.D	14 Apr 2025 15:07		AR\AJ	Ok,M
6	PSTDICC075	PSTDICC075	PL095206.D	14 Apr 2025 15:21		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL095207.D	14 Apr 2025 15:35		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL095208.D	14 Apr 2025 16:02		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PL095209.D	14 Apr 2025 16:15		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL095210.D	14 Apr 2025 16:29		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL095211.D	14 Apr 2025 16:43		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL095212.D	14 Apr 2025 16:56		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL095213.D	14 Apr 2025 17:10		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL095214.D	14 Apr 2025 17:24		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PL095215.D	14 Apr 2025 17:38		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL095216.D	14 Apr 2025 17:51		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL095217.D	14 Apr 2025 18:05		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL095218.D	14 Apr 2025 18:19		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL041425

Review By	Abdul	Review On	4/15/2025 7:51:14 AM
Supervise By	mohammad	Supervise On	4/16/2025 1:08:48 AM
SubDirectory	PL041425	HP Acquire Method	HP Processing Method pl041425 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL095219.D	14 Apr 2025 18:32		AR\AJ	Ok
20	PSTDICV050	ICVPL041425	PL095220.D	14 Apr 2025 18:46		AR\AJ	Ok
21	PCHLORICV500	ICVPL041425CHLOR	PL095221.D	14 Apr 2025 19:00		AR\AJ	Ok
22	PTOXICV500	ICVPL041425TOX	PL095222.D	14 Apr 2025 19:13		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL041525

Review By	Abdul	Review On	4/16/2025 8:41:15 AM
Supervise By	mohammad	Supervise On	4/16/2025 9:16:31 AM
SubDirectory	PL041525	HP Acquire Method	HP Processing Method p1041425 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24095
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL095223.D	15 Apr 2025 10:24		AR\AJ	Ok
2	I.BLK	I.BLK	PL095224.D	15 Apr 2025 10:37		AR\AJ	Ok
3	PEM	PEM	PL095225.D	15 Apr 2025 10:51		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL095226.D	15 Apr 2025 11:21		AR\AJ	Ok,M
5	Q1784-01	OR-01-41125	PL095227.D	15 Apr 2025 11:35		AR\AJ	Ok,M
6	Q1783-01	SU-03-41125	PL095228.D	15 Apr 2025 11:48		AR\AJ	Ok,M
7	Q1792-01	TR-05-041125	PL095229.D	15 Apr 2025 12:02		AR\AJ	Ok,M
8	PB167576BL	PB167576BL	PL095230.D	15 Apr 2025 12:16		AR\AJ	Ok
9	PB167576BS	PB167576BS	PL095231.D	15 Apr 2025 12:29	F Flag coming	AR\AJ	Not Ok
10	PB167535BS	PB167535BS	PL095232.D	15 Apr 2025 13:05		AR\AJ	Ok,M
11	Q1785-01	HR-02-041125	PL095233.D	15 Apr 2025 13:30		AR\AJ	Ok,M
12	Q1785-03	HR-03-041125	PL095234.D	15 Apr 2025 13:43		AR\AJ	Ok,M
13	I.BLK	I.BLK	PL095235.D	15 Apr 2025 13:57		AR\AJ	Ok
14	PSTDCCC050	PSTDCCC050	PL095236.D	15 Apr 2025 15:08		AR\AJ	Ok,M
15	PB167576BS	PB167576BS	PL095237.D	15 Apr 2025 15:22		AR\AJ	Ok,M
16	Q1787-09	TP-1	PL095238.D	15 Apr 2025 15:40		AR\AJ	Ok,M
17	Q1787-09MS	TP-1MS	PL095239.D	15 Apr 2025 15:53		AR\AJ	Ok,M
18	Q1787-09MSD	TP-1MSD	PL095240.D	15 Apr 2025 16:07		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL041525

Review By	Abdul	Review On	4/16/2025 8:41:15 AM
Supervise By	mohammad	Supervise On	4/16/2025 9:16:31 AM
SubDirectory	PL041525	HP Acquire Method	HP Processing Method pl041425 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1788-01	WC-1	PL095241.D	15 Apr 2025 16:21		AR\AJ	Ok,M
20	PB167596BL	PB167596BL	PL095242.D	15 Apr 2025 16:35		AR\AJ	Ok
21	PB167596BS	PB167596BS	PL095243.D	15 Apr 2025 16:48		AR\AJ	Ok,M
22	Q1806-01	OR-03-041425	PL095244.D	15 Apr 2025 17:02		AR\AJ	Ok,M
23	Q1807-01	OK-02-041425	PL095245.D	15 Apr 2025 17:16		AR\AJ	Ok,M
24	Q1808-01	OILY-SOIL-PILE	PL095246.D	15 Apr 2025 17:30		AR\AJ	Ok,M
25	Q1808-01MS	OILY-SOIL-PILEMS	PL095247.D	15 Apr 2025 17:43		AR\AJ	Ok,M
26	Q1808-01MSD	OILY-SOIL-PILEMSD	PL095248.D	15 Apr 2025 17:57		AR\AJ	Ok,M
27	Q1808-03	LAW-25-0060	PL095249.D	15 Apr 2025 18:10		AR\AJ	Ok,M
28	I.BLK	I.BLK	PL095250.D	15 Apr 2025 18:38		AR\AJ	Ok
29	PEM	PEM	PL095251.D	15 Apr 2025 18:52		AR\AJ	Ok,M
30	PSTDCCC050	PSTDCCC050	PL095252.D	15 Apr 2025 19:05		AR\AJ	Ok,M

M : Manual Integration