

Surrogate Summary

SDG No.: Q1739

Client: PARSONS Engineering of New York, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL094566.D	PIBLK-PL094566.D	Decachlorobiphenyl	1	20	22.7	114		43	140
		Tetrachloro-m-xylene	1	20	20.6	103		77	126
		Decachlorobiphenyl	2	20	21.0	105		43	140
		Tetrachloro-m-xylene	2	20	20.2	101		77	126
I.BLK-PL095137.D	PIBLK-PL095137.D	Decachlorobiphenyl	1	20	20.9	105		43	140
		Tetrachloro-m-xylene	1	20	19.1	96		77	126
		Decachlorobiphenyl	2	20	16.8	84		43	140
		Tetrachloro-m-xylene	2	20	20.4	102		77	126
Q1739-02	WC-LIQUID-20250404	Decachlorobiphenyl	1	20	17.2	86		43	140
		Tetrachloro-m-xylene	1	20	11.7	58	*	77	126
		Decachlorobiphenyl	2	20	12.6	63		43	140
		Tetrachloro-m-xylene	2	20	12.7	63	*	77	126
Q1739-02MS	WC-LIQUID-20250404MS	Decachlorobiphenyl	1	20	17.4	87		43	140
		Tetrachloro-m-xylene	1	20	12.5	62	*	77	126
		Decachlorobiphenyl	2	20	12.8	64		43	140
		Tetrachloro-m-xylene	2	20	14.0	70	*	77	126
Q1739-02MSD	WC-LIQUID-20250404MSD	Decachlorobiphenyl	1	20	16.4	82		43	140
		Tetrachloro-m-xylene	1	20	12.3	61	*	77	126
		Decachlorobiphenyl	2	20	12.9	64		43	140
		Tetrachloro-m-xylene	2	20	13.7	69	*	77	126
I.BLK-PL095147.D	PIBLK-PL095147.D	Decachlorobiphenyl	1	20	18.8	94		43	140
		Tetrachloro-m-xylene	1	20	18.1	90		77	126
		Decachlorobiphenyl	2	20	15.8	79		43	140
		Tetrachloro-m-xylene	2	20	19.1	95		77	126
PB167535BL	PB167535BL	Decachlorobiphenyl	1	20	20.9	104		43	140
		Tetrachloro-m-xylene	1	20	19.3	97		77	126
		Decachlorobiphenyl	2	20	19.5	98		43	140
		Tetrachloro-m-xylene	2	20	20.2	101		77	126
PB167488TB	PB167488TB	Decachlorobiphenyl	1	20	20.8	104		43	140
		Tetrachloro-m-xylene	1	20	18.9	95		77	126
		Decachlorobiphenyl	2	20	20.2	101		43	140
		Tetrachloro-m-xylene	2	20	19.6	98		77	126
I.BLK-PL095162.D	PIBLK-PL095162.D	Decachlorobiphenyl	1	20	22.4	112		43	140
		Tetrachloro-m-xylene	1	20	19.1	96		77	126
		Decachlorobiphenyl	2	20	22.0	110		43	140
		Tetrachloro-m-xylene	2	20	20.1	100		77	126
I.BLK-PL095202.D	PIBLK-PL095202.D	Decachlorobiphenyl	1	20	22.6	113		43	140
		Tetrachloro-m-xylene	1	20	20.6	103		77	126
		Decachlorobiphenyl	2	20	21.7	109		43	140
		Tetrachloro-m-xylene	2	20	20.1	100		77	126
I.BLK-PL095224.D	PIBLK-PL095224.D	Decachlorobiphenyl	1	20	23.9	120		43	140

Surrogate Summary

SDG No.: Q1739

Client: PARSONS Engineering of New York, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL095224.D	PIBLK-PL095224.D	Tetrachloro-m-xylene	1	20	22.7	114		77	126
		Decachlorobiphenyl	2	20	22.3	111		43	140
PB167535BS	PB167535BS	Tetrachloro-m-xylene	2	20	20.2	101		77	126
		Decachlorobiphenyl	1	20	19.0	95		43	140
		Tetrachloro-m-xylene	1	20	16.6	83		77	126
		Decachlorobiphenyl	2	20	15.4	77		43	140
		Tetrachloro-m-xylene	2	20	15.3	77		77	126
		Decachlorobiphenyl	1	20	20.9	105		43	140
I.BLK-PL095235.D	PIBLK-PL095235.D	Tetrachloro-m-xylene	1	20	23.8	119		77	126
		Decachlorobiphenyl	2	20	15.2	76		43	140
		Tetrachloro-m-xylene	2	20	20.7	103		77	126
		Decachlorobiphenyl	1	20	20.9	105		43	140