

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1739

Client: PARSONS Engineering of New York, In

Analytical Method: **8081B** **Datafile :** PL095232.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167535BS	gamma-BHC (Lindane)	0.5	0.46	ug/L	92				82	129	
	Heptachlor	0.5	0.50	ug/L	100				79	127	
	Heptachlor epoxide	0.5	0.49	ug/L	98				81	124	
	Endrin	0.5	0.52	ug/L	105				81	128	
	Methoxychlor	0.5	0.52	ug/L	104				78	108	