

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: PARS02
ProjectID: Con Edison - 11th Ave-West 50th St Site
Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG No.: Q1739
DataFile: FF015773.D Analyst Name: YP\AJ Analyst Date: 04-09-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	95564124	112428	110052	2.159