

WATER TPH GC LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** PARSONS Engineering of New York, Inc.
Lab Code: CHEM **Cas No:** Q1739 **SAS No :** Q1739 **SDG No:** Q1739
Matrix Spike - EPA Sample No : PB167520BS **Datafile:** FF015770.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS
Petroleum Hydrocarbons	340	0	267	78	78-117