

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM040825

Review By	anahy	Review On	4/9/2025 10:38:14 AM		
Supervise By	Jagrut	Supervise On	4/9/2025 4:50:06 PM		
SubDirectory	BM040825	HP Acquire Method	BNA_M	HP Processing Method	bm040825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12657,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM049847.D	08 Apr 2025 12:55		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM049848.D	08 Apr 2025 13:35		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BM049849.D	08 Apr 2025 14:14	Compound #32,54,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BM049850.D	08 Apr 2025 14:53	Compound #32,54 kept on LR	RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM049851.D	08 Apr 2025 15:32	This calibration is fail for Com#77	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BM049852.D	08 Apr 2025 16:12	This calibration is good for 8270E DOD except Com#77 and good for 625.1 method	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BM049853.D	08 Apr 2025 17:30		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BM049854.D	08 Apr 2025 19:28	Compound #69 Removed from 60PPM	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BM049855.D	08 Apr 2025 20:07	Compound #69 Removed from 80PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBM040825	BM049856.D	08 Apr 2025 20:47		RC/JU	Ok
11	PB167474BL	PB167474BL	BM049857.D	08 Apr 2025 21:26	Use for Q1730 only	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM040925

Review By	anahy	Review On	4/10/2025 10:38:09 AM
Supervise By	Jagrut	Supervise On	4/10/2025 3:33:56 PM
SubDirectory	BM040925	HP Acquire Method	BNA_M
		HP Processing Method	bm040825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12658,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM049858.D	09 Apr 2025 08:30		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM049859.D	09 Apr 2025 09:09		RC/JU	Ok
3	PB167521BL	PB167521BL	BM049860.D	09 Apr 2025 09:48		RC/JU	Ok
4	PB167521BS	PB167521BS	BM049861.D	09 Apr 2025 10:27		RC/JU	Ok
5	PB167521BSD	PB167521BSD	BM049862.D	09 Apr 2025 11:06		RC/JU	Ok
6	Q1740-01	TP-20	BM049863.D	09 Apr 2025 11:50		RC/JU	Ok
7	Q1744-01	B-158-SB01	BM049864.D	09 Apr 2025 12:29		RC/JU	Ok
8	Q1744-01MS	B-158-SB01MS	BM049865.D	09 Apr 2025 13:08		RC/JU	Ok
9	Q1744-01MSD	B-158-SB01MSD	BM049866.D	09 Apr 2025 13:47		RC/JU	Ok
10	Q1742-01	TR-06-040725	BM049867.D	09 Apr 2025 14:26		RC/JU	Ok,M
11	Q1746-03	B-149-SB02	BM049868.D	09 Apr 2025 15:05		RC/JU	Ok
12	Q1744-03	B-158-SB02	BM049869.D	09 Apr 2025 15:45		RC/JU	Ok
13	Q1712-01DL	Z-05ADL	BM049870.D	09 Apr 2025 16:24	Need 5X further Dilution	RC/JU	Dilution
14	Q1712-01DL2	Z-05ADL2	BM049871.D	09 Apr 2025 17:03		RC/JU	Ok,M
15	Q1743-01	TP-16	BM049872.D	09 Apr 2025 17:42		RC/JU	Ok,M
16	Q1745-09	IB-6.5-WC	BM049873.D	09 Apr 2025 18:21		RC/JU	Ok,M
17	Q1739-01	WC-LIQUID-20250404	BM049874.D	09 Apr 2025 19:01	Surrogate Fail	RC/JU	Ok

M : Manual Integration