

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	04/02/25	
Project:	Con Edison - 11th Ave-West 50th St Site		Date Received:	04/03/25	
Client Sample ID:	Z-05AMSD		SDG No.:	Q1739	
Lab Sample ID:	Q1712-04MSD		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024236.D	1	04/08/25 08:25	04/09/25 15:39	PB167518

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	360		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	450		5.30	50.0	ug/L
95-48-7	2-Methylphenol	510		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	500		11.0	100	ug/L
67-72-1	Hexachloroethane	450		6.50	50.0	ug/L
98-95-3	Nitrobenzene	490		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	450		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	540		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	550		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	590		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	510		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	1100	E	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	140		10 - 139	93%	SPK: 150
13127-88-3	Phenol-d6	126		10 - 134	84%	SPK: 150
4165-60-0	Nitrobenzene-d5	99.4		49 - 133	99%	SPK: 100
321-60-8	2-Fluorobiphenyl	92.2		52 - 132	92%	SPK: 100
118-79-6	2,4,6-Tribromophenol	160		44 - 137	107%	SPK: 150
1718-51-0	Terphenyl-d14	101		48 - 125	101%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	294000		7.74		
1146-65-2	Naphthalene-d8	1160000		10.516		
15067-26-2	Acenaphthene-d10	692000		14.375		
1517-22-2	Phenanthrene-d10	1330000		17.181		
1719-03-5	Chrysene-d12	1360000		21.633		
1520-96-3	Perylene-d12	1480000		25.004		

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	04/02/25	
Project:	Con Edison - 11th Ave-West 50th St Site		Date Received:	04/03/25	
Client Sample ID:	Z-05AMSD		SDG No.:	Q1739	
Lab Sample ID:	Q1712-04MSD		Matrix:	TCLP	
Analytical Method:	SW8270		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024236.D	1	04/08/25 08:25	04/09/25 15:39	PB167518

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products