

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1739

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1712-04MS	Client Sample ID:	Z-05AMS					DataFile:	BP024235.D		
Pyridine	500	0	350	ug/L	70				10	109	
1,4-Dichlorobenzene	500	0	420	ug/L	84				55	125	
2-Methylphenol	500	0	490	ug/L	98				60	131	
3+4-Methylphenols	500	0	490	ug/L	98				54	136	
Hexachloroethane	500	0	430	ug/L	86				19	146	
Nitrobenzene	500	0	470	ug/L	94				62	112	
Hexachlorobutadiene	500	0	420	ug/L	84				52	125	
2,4,6-Trichlorophenol	500	0	490	ug/L	98				78	112	
2,4,5-Trichlorophenol	500	0	500	ug/L	100				71	111	
2,4-Dinitrotoluene	500	0	570	ug/L	114				74	137	
Hexachlorobenzene	500	0	470	ug/L	94				72	115	
Pentachlorophenol	1000	0	1100	ug/L	110				52	162	