

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1739

Client: PARSONS Engineering of New York, Inc.

Analytical Method: 8270E

DataFile: BP024263.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB167518BS	Pyridine	50	42.0	ug/L	84				29	97	
	1,4-Dichlorobenzene	50	50.6	ug/L	101				76	103	
	2-Methylphenol	50	57.5	ug/L	115		*		69	109	
	3+4-Methylphenols	50	56.1	ug/L	112		*		67	106	
	Hexachloroethane	50	51.8	ug/L	104				76	118	
	Nitrobenzene	50	52.0	ug/L	104				58	106	
	Hexachlorobutadiene	50	51.1	ug/L	102		*		69	101	
	2,4,6-Trichlorophenol	50	55.7	ug/L	111		*		61	110	
	2,4,5-Trichlorophenol	50	55.0	ug/L	110		*		70	106	
	2,4-Dinitrotoluene	50	58.7	ug/L	117		*		60	115	
	Hexachlorobenzene	50	54.0	ug/L	108		*		73	106	
	Pentachlorophenol	100	120	ug/L	120		*		47	114	