

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG NO.: Q1739

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/10/2025

Lab File ID: BP024245.D Time Analyzed: 09:46

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	322953	7.746	1317780	10.53	792808	14.39
UPPER LIMIT	645906	8.246	2635560	11.028	1585620	14.892
LOWER LIMIT	161477	7.246	658890	10.028	396404	13.892
EPA SAMPLE NO.						
01 PB167518BL	279607	7.75	1127550	10.53	674843	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.