

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG NO.: Q1739

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/11/2025

Lab File ID: BP024259.D Time Analyzed: 09:43

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	312531	7.757	1264670	10.53	754959	14.39
UPPER LIMIT	625062	8.257	2529340	11.034	1509920	14.892
LOWER LIMIT	156266	7.257	632335	10.034	377480	13.892
EPA SAMPLE NO.						
01 PB167518BS	270274	7.76	1101810	10.54	656088	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.