

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG NO.: Q1739

EPA Sample No.: SSTDICCC040 Date Analyzed: 04/14/2025

Lab File ID: BP024279.D Time Analyzed: 13:49

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	246363	7.734	1046880	10.51	657357	14.36
UPPER LIMIT	492726	8.234	2093760	11.005	1314710	14.863
LOWER LIMIT	123182	7.234	523440	10.005	328679	13.863
EPA SAMPLE NO.						
01 PB167488TB	225183	7.73	887369	10.50	527513	14.36

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.