

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PARS02

Lab Code: CHEM

SAS No.: Q1739

SDG NO.: Q1739

Lab File ID: BP024227.D

DFTPP Injection Date: 04/09/2025

Instrument ID: BNA_P

DFTPP Injection Time: 08:42

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	28.2
68	Less than 2.0% of mass 69	0.6 (1.6) 1
69	Mass 69 relative abundance	34.3
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	46.7
197	Less than 2.0% of mass 198	0.5
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	28.9
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	14.5
442	Greater than 50% of mass 198	92.9
443	15.0 - 24.0% of mass 442	18.3 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024228.D	04/09/2025	10:08
Z-05AMS	Q1712-04MS	BP024235.D	04/09/2025	14:59
Z-05AMSD	Q1712-04MSD	BP024236.D	04/09/2025	15:39