

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PARS02

Lab Code: CHEM

SAS No.: Q1739

SDG NO.: Q1739

Lab File ID: BP024244.D

DFTPP Injection Date: 04/10/2025

Instrument ID: BNA_P

DFTPP Injection Time: 09:05

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	28.8
68	Less than 2.0% of mass 69	0.6 (1.7) 1
69	Mass 69 relative abundance	34.8
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	47
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	28.5
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	14.6
442	Greater than 50% of mass 198	95.2
443	15.0 - 24.0% of mass 442	18.8 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024245.D	04/10/2025	09:46
PB167518BL	PB167518BL	BP024246.D	04/10/2025	10:27