

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG No.: Q1739

Instrument ID: MSVOA_X Calibration Date/Time: 04/09/2025 10:35

Lab File ID: VX045664.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.703	0.690		-1.85	20
1,1-Dichloroethene	0.600	0.586		-2.33	20
2-Butanone	0.550	0.538		-2.18	20
Carbon Tetrachloride	0.518	0.553		6.76	20
Chloroform	1.306	1.317		0.84	20
Benzene	1.463	1.461		-0.14	20
1,2-Dichloroethane	0.604	0.626		3.64	20
Trichloroethene	0.348	0.351		0.86	20
Tetrachloroethene	0.353	0.373		5.67	20
Chlorobenzene	1.068	1.118	0.3	4.68	20
1,2-Dichloroethane-d4	0.914	0.945		3.39	20
Dibromofluoromethane	0.355	0.379		6.76	20
Toluene-d8	1.238	1.272		2.75	20
4-Bromofluorobenzene	0.451	0.501		11.09	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.