

### Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Edison - 11th Ave-West 50th St Site		Date Received:	
Client Sample ID:	VX0409WBSD01		SDG No.:	Q1739
Lab Sample ID:	VX0409WBSD01		Matrix:	TCLP
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	TCLP VOA
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045668.D	1		04/09/25 12:14	VX040925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-01-4	Vinyl Chloride	18.9		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.2		0.23	1.00	ug/L
78-93-3	2-Butanone	100		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	21.2		0.25	1.00	ug/L
67-66-3	Chloroform	20.8		0.25	1.00	ug/L
71-43-2	Benzene	20.4		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	21.5		0.22	1.00	ug/L
79-01-6	Trichloroethene	20.4		0.090	1.00	ug/L
127-18-4	Tetrachloroethene	22.0		0.23	1.00	ug/L
108-90-7	Chlorobenzene	20.9		0.12	1.00	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	54.1		74 - 125	108%	SPK: 50
1868-53-7	Dibromofluoromethane	53.8		75 - 124	108%	SPK: 50
2037-26-5	Toluene-d8	53.1		86 - 113	106%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.3		77 - 121	109%	SPK: 50
<b>INTERNAL STANDARDS</b>						
363-72-4	Pentafluorobenzene	83900	5.544			
540-36-3	1,4-Difluorobenzene	147000	6.757			
3114-55-4	Chlorobenzene-d5	129000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	60500	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products