

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q1739

**Client:** PARSONS Engineering of New York, Inc.

**Analytical Method:** SW8260-Low

**Datafile :** VX045667.D

| Lab Sample ID | Parameter            | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits<br>High | RPD |
|---------------|----------------------|-------|--------|------|-----|-----|------|-----|----------------|-----|
| VX0409WBS01   | Vinyl chloride       | 20    | 18.5   | ug/L | 93  |     |      | 65  | 117            |     |
|               | 1,1-Dichloroethene   | 20    | 18.4   | ug/L | 92  |     |      | 74  | 110            |     |
|               | 2-Butanone           | 100   | 94.6   | ug/L | 95  |     |      | 65  | 122            |     |
|               | Carbon Tetrachloride | 20    | 20.1   | ug/L | 101 |     |      | 77  | 113            |     |
|               | Chloroform           | 20    | 19.3   | ug/L | 97  |     |      | 79  | 113            |     |
|               | Benzene              | 20    | 19.3   | ug/L | 97  |     |      | 82  | 109            |     |
|               | 1,2-Dichloroethane   | 20    | 20.4   | ug/L | 102 |     |      | 80  | 115            |     |
|               | Trichloroethene      | 20    | 19.2   | ug/L | 96  |     |      | 77  | 113            |     |
|               | Tetrachloroethene    | 20    | 20.9   | ug/L | 104 |     |      | 67  | 123            |     |
|               | Chlorobenzene        | 20    | 20.3   | ug/L | 102 |     |      | 82  | 109            |     |