

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG NO.: Q1739
 Lab File ID: VX045664.D Date Analyzed: 04/09/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:35
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	98594	5.54	170453	6.75	147839	10.05
UPPER LIMIT	197188	6.044	340906	7.251	295678	10.549
LOWER LIMIT	49297	5.044	85226.5	6.251	73919.5	9.549
EPA SAMPLE NO.						
WC-LIQUID-20250404	64975	5.54	128842	6.76	121333	10.05
VX0409WBL01	69921	5.54	138334	6.76	127856	10.05
VX0409WBS01	94074	5.54	163361	6.76	141398	10.05
VX0409WBSD01	83922	5.54	147335	6.76	129399	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.