

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: Q1739 SAS No.: Q1739 SDG NO.: Q1739
 Lab File ID: VX045664.D Date Analyzed: 04/09/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:35
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	71972	12.018				
UPPER LIMIT	143944	12.518				
LOWER LIMIT	35986	11.518				
EPA SAMPLE NO.						
WC-LIQUID-20250404	54281	12.02				
VX0409WBL01	50474	12.02				
VX0409WBS01	64175	12.02				
VX0409WBSD01	60456	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.