

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1741 SAS No.: Q1741 SDG NO.: Q1741

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 04/11/2025

Lab File ID: BN036884.D Time Analyzed: 18:45

Instrument ID: BNA\_N GC Column: ZB-GR ID: 0.25 (mm)

|                      | IS1 (DCB)<br>AREA # | RT # | IS2 (NPT)<br>AREA # | RT #   | IS3 (ANT)<br>AREA # | RT #   |
|----------------------|---------------------|------|---------------------|--------|---------------------|--------|
| 12 HOUR STD          | 254                 | 7.64 | 679                 | 10.43  | 407                 | 14.29  |
| UPPER LIMIT          | 508                 | 8.14 | 1358                | 10.925 | 814                 | 14.788 |
| LOWER LIMIT          | 127                 | 7.14 | 339.5               | 9.925  | 203.5               | 13.788 |
| EPA SAMPLE NO.       |                     |      |                     |        |                     |        |
| 01 PB167519BS        | 385                 | 7.64 | 924                 | 10.43  | 537                 | 14.29  |
| 02 PB167519BSD       | 195                 | 7.64 | 480                 | 10.43  | 292                 | 14.29  |
| 03 RMW-02B-66-040425 | 274                 | 7.64 | 725                 | 10.43  | 456                 | 14.29  |
| 04 RMW-06B-74-040425 | 257                 | 7.64 | 678                 | 10.43  | 429                 | 14.29  |
| 05 FB01-040425       | 267                 | 7.64 | 721                 | 10.43  | 449                 | 14.29  |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.