

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1741 SAS No.: Q1741 SDG NO.: Q1741

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 04/14/2025

Lab File ID: BN036900.D Time Analyzed: 14:38

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1387	7.633	3400	10.42	1899	14.28
UPPER LIMIT	2774	8.133	6800	10.915	3798	14.777
LOWER LIMIT	693.5	7.133	1700	9.915	949.5	13.777
EPA SAMPLE NO.						
01 PB167519BL	1898	7.63	4409	10.42	2284	14.28
02 RMW-02B-66-040425DL	1215	7.63	2952	10.42	1740	14.28

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.