

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1741

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Datafile : VX045618.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0407WBSD01	Vinyl chloride	20	18.6	ug/L	93	3		70 (65)	130 (117)	20 (20)
	1,1-Dichloroethene	20	18.9	ug/L	95	8		70 (74)	130 (110)	20 (20)
	1,1-Dichloroethane	20	18.7	ug/L	94	5		70 (78)	130 (112)	20 (20)
	cis-1,2-Dichloroethene	20	18.6	ug/L	93	6		70 (77)	130 (110)	20 (20)
	1,1,1-Trichloroethane	20	19.2	ug/L	96	6		70 (80)	130 (108)	20 (20)
	Benzene	20	19.2	ug/L	96	6		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	20.3	ug/L	102	8		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	19.2	ug/L	96	6		70 (77)	130 (113)	20 (20)
	1,1,2-Trichloroethane	20	20.2	ug/L	101	10		70 (83)	130 (112)	20 (20)
	Tetrachloroethene	20	20.1	ug/L	101	4		70 (67)	130 (123)	20 (20)