

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05  
 Lab Code: CHEM Case No.: Q1741 SAS No.: Q1741 SDG NO.: Q1741  
 Lab File ID: VX045614.D Date Analyzed: 04/07/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 09:17  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                   | IS4<br>AREA # | RT #   |  |  |  |  |
|-------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD       | 73594         | 12.018 |  |  |  |  |
| UPPER LIMIT       | 147188        | 12.518 |  |  |  |  |
| LOWER LIMIT       | 36797         | 11.518 |  |  |  |  |
| EPA SAMPLE NO.    |               |        |  |  |  |  |
| RMW-06B-74-040425 | 44488         | 12.02  |  |  |  |  |
| FB01-040425       | 49856         | 12.02  |  |  |  |  |
| TB01-040425       | 47648         | 12.02  |  |  |  |  |
| VX0407WBL01       | 49926         | 12.02  |  |  |  |  |
| VX0407WBS01       | 64861         | 12.02  |  |  |  |  |
| VX0407WBSD01      | 67274         | 12.02  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.