

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture
Project: Amtrak Sawtooth Bridges 2025

SDG No.: Q1744
RunNo.: LB135327

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.01	7	100	90-110	04/07/2025
Sample ID: Corrosivity	CCV1	pH	2.02	2.00	101	90-110	04/07/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	04/07/2025

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: Q1744

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB135328

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	04/07/2025
Sample ID:	CCV1						
pH		pH	2.02	2.00	101	90-110	04/07/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	04/07/2025

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: Q1744

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB135331

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.495	0.5	99	90-110	04/07/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.484	0.5	97	90-110	04/07/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.489	0.5	98	90-110	04/07/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.485	0.5	97	90-110	04/07/2025

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: Q1744

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB135341

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.09	0.099	91	85-115	04/08/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	04/08/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	04/08/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	04/08/2025